



Valley Safety Services Associates, Inc.

330 Old Enfield Road, Belchertown, MA 01007

<http://www.vssa-inc.com>

Tel/Fax: (413)323-9571

Leak Test Data/Certificate

Complete the data requested on the top part of this form and return it with your leak test to the address in the above heading. Following our analysis of your leak test specimen we will return this form to you for your record of the test and results.

Source/Device Description

Device: _____ Model: _____ Serial #: _____ Manufacturer: _____

Radioisotope _____ Activity (mCi) _____

Source model: _____ Source #: _____ Manufacturer: _____

The leak testing of this source/device was performed as indicated in manufacturer recommendations or LT-952 kit instructions using kit # _____

By: _____ Date: _____

Company: _____ Telephone: _____

Address: _____ Email: _____

Do Not Write Below This Line

Leak Test Analysis Result

Analysis of the above test on _____ yielded the following;

- Statistical analysis of the radioactive count data of this leak test specimen indicated any activity present is less than 0.005 μ Ci. The source may be used as authorized.
- Statistical analysis of the radioactive count data of this leak test specimen indicated there is greater than 0.005 microcuries of activity present. This source should be considered leaking. Consult your device operations manual; place this unit in storage and make the required notification to your regulatory agency.

REMINDER: YOUR NEXT REQUIRED LEAK TEST FOR THIS DEVICE/SOURCE IS DUE ON OR BEFORE

ANOTHER TEST KIT WILL BE SENT TO YOU 2 WEEKS BEFORE THIS DATE FOR YOUR USE AND RETURN FOR ANALYSIS. (Please ensure to check for the requirements of your state.)

THIS CERTIFICATE IS AN ESSENTIAL RECORD AND SHOULD BE MAINTAINED FOR INSPECTION BY THE REGULATORY AGENCY.

CERTIFICATE #: _____ BY: _____ DATE: _____